


| | |
|-------------------------------------|------------------------------------|
| Pressure-Rating of the sensor probe | As specified in order confirmation |
| Material | 316L |
| Process connection | Tri-Clamp 1.5" |
| Compatible with | ISO1127 DN 25 / 32 |
| | DIN 32676 DN 25* / 32 / 40 |
| | ASME-BPE 1.5" |
| Order Code | SRV-X3-15T |

*ID of the ferrule should be at least 26mm for the sensor to fit, check prior to ordering



Units: mm
[inch]

SRV-X3-15T Tri Clamp 1.5"

This drawing remains property of rheonics. It may be neither copied nor informed to third persons or competitors without our approval. (2 § paragraph 1 fig.7 copyright act of 9th Sept. 1965)

Protected by US and International patents granted and pending



inline process
density and viscosity
monitoring